

ABLE FT FLAREMASTER METERING SYSTEM SIL2 CERTIFICATE

FMEDA REPORT 01-05-005-1.A

issued by

ABLE Instruments & Controls Ltd

Herewith we confirm ABLE FT flare meter is SIL2 certified in accord with IEC 61508-4 and can be applied in safety related applications. When correct pressure and temperature Transmitters are applied, ABLE FT FlareMaster systems fulfil the requirements for SIL2. ABLE FT flare meters can also fulfil the requirements for SIL4 with redundant architecture.



Replication of certificate issued by Michat Bigus
Certified Functional Safety Specialist UTD No. 11-1-90219
Member of Functional Safety Forum

Certyfikowany Specjalista ds. Bezpieczeństwa funkcjonalnego Michat Bigus